

Understand a therapeutic adherence report

Compliance Summary

Date Range	11/16/2021 - 12/15/2021 (30 days)
Days with Device Usage	30 days
Average Usage (All Days)	7 hrs. 51 mins. 4 secs.
1 Average Usage (Days Used)	7 hrs. 51 mins. 4 secs.
Minimum Usage (1 Day)	5 hrs. 26 mins. 28 secs.

CPAP Summary

Average Time in Large Leak Per Day	4 mins. 46 secs.
2 Average AHI	0.5
CPAP	16.0 cmH2O


Device Settings as of 12/15/2021

Device Mode CPAP - C-Flex+

Device Settings

Parameter	Value
Auto-Trial	Off
CPAP Pressure	16 cmH2O
C-Flex+ Setting	3
C-Flex+ Lock	Off

Percent of Night in Large Leak (LL)



Need Support?

Visit our [new support section](#) for instructional videos, user guides and a host of other interesting information.

Please note that the data provided are for information purposes only, they do not represent your personal data.

1 To benefit from the positive effects of your therapy, the hours of treatment should be as close as possible to your sleeping time. We encourage you to wear your equipment every night.

2 In order for your therapy to be optimal, the Index of Apnea and Hypopnea (IAH) should be less than 10 per hour. If this is not your case, we invite you to discuss about it with your respiratory therapist.

3 The limit on the acceptable leakage value is 60 l/min. If you do not reach the goal:

Make sure that the fit of your mask is adequate

Improper maintenance and wear of your mask may also cause leaks

Questions?

Contact us at 1 866-341-4975